
Technical Specifications

AT235

Impedance Audiometer



License Overview

International configuration AT235		
	Standard	Standard + H
Licenses Tympanometry		
Test:		
Tymp 226Hz	x	x
Tymp 678, 800 & 1000Hz		x
Manual tympanometry		x
Reflex Ipsilateral	x	x
Reflex Contralateral	x	x
ETF 1 test non-perforated tympanic membrane	x	x
EFT 2 test perforated tympanic membrane		x
EFT 3 patulous tympanic membrane		x
Reflex Decay	x	x
Reflex Latency		x
Pure tone audiometry	x	x
Modified Hughson Westlake audiometry test	x	x
Tympanogram transfer	Optional	Optional

Languages supported in IMP and AUD

	Chinese	Czech	English	Finnish	French	German	Greek	Italian	Japanese	Korean	Norwegian	Polish	Portuguese	Russian	Spanish	Turkish
IMP																
MT10			X		X	X										
Existing AT235			X			X										
New AT235	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X
Existing AA222			X			X										
New AA222	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X
Titan	X	X	X		X	X		X	X	X			X	X	X	
AUD																
AS608			X		X	X									X	
AD226	X		X		X	X		X				X	X	X	X	X
AD629 / AD229	X *	X *	X	X *	X	X	X *	X	X*	X *	X *	X	X *	X *	X	X
AC40	X *	X *	X	X *	X	X	X *	X	X*	X *	X *	X	X *	X *	X	X
Suites																
Titan suite	X	X	X		X	X		X	X			X	X	X	X	X
DS	X	X	X		X	X	X	X	X	X		X	X	X	X	X

Included and Optional Parts

Included Parts:

AT235 instrument

Clinical Probe System¹

Diagnostic Probe System¹

Power supply unit UE60

Contralateral headphone¹

Audiometric Headset¹

CAT50

Patient Response¹

Printer kit (Options)

Wall mounting kit (Options)

Cleaning cloth

Assortment bag BET55

¹ Applied part as according to IEC60601-1

General Technical Specifications

AT235 Technical Specifications

Medical CE-mark:	The CE-mark indicates that Interacoustics A/S meets the requirements of Annex II of the Medical Device Directive 93/42/EEC. Approval of the quality system is made by TÜV – identification no0123	
Standards:	Safety:	IEC 60601-1, Class I, Type B applied parts
	EMC:	IEC 60601-1-2
	Impedance:	IEC 60645-5/ANSI S3.39, Type 1
	Audiometer:	IEC60645-1/ANSI S3.6, Type 4
Operation environment:	Temperature:	15 – 35 °C
	Relative Humidity:	30 – 90%
	Ambient Pressure:	98kPa – 104kPa
	Warm-up Time:	1 minute
Transport & Storage:	Storage Temperature:	0°C – 50°C
	Transport Temperature:	-20 – 50 °C
	Rel. Humidity:	10 – 95%
Internal Battery	CR2032 3V, 230mAh, Li. Not serviceable by user.	
Impedance Measuring System		
Probe tone:	Frequency:	226 Hz, 678 Hz, 800 Hz, 1000 Hz; pure tones; ±1%
	Level:	85 dB SPL (\approx 69 dB HL) \pm 1.5 dB
Air pressure:	Control:	Automatic.
	Indicator:	Measured value is displayed on the graphical display.
	Range:	-600 to +400 daPa. \pm 5%
	Pressure limitation:	-750 daPa and +550 daPa.
Compliance:	Range:	0.1 to 8.0 ml at 226 Hz probe tone (Ear volume: 0.1 to 8.0 ml) and 0.1 to 15 mmho at 678, 800 and 1000 Hz probe tone. All \pm 5%
Test types:	Tympanometry	Automatic, where the start and stop pressure can be user-programmed in the setup function.
		Manual control of all functions.
	Eustachian tube function 1 – Non perforated eardrum	Williams test
	Eustachian tube function 2 – Perforated eardrum	Toynbee test
	Eustachian tube function 3 – Patulous Eustachian tube	Continuous sensitive impedance measurement


Reflex Functions		
Signal sources:	Tone - Contra, Reflex:	250, 500, 1000, 2000, 3000, 4000, 6000, 8000 Hz, Wide Band, High and Low pass
	Tone - Ipsi, Reflex:	500, 1000, 2000, 3000, 4000 Hz wide band, high and low pass.
	NB noise – Contra, Reflex	250, 500, 1000, 2000, 3000, 4000, 6000, 8000 Hz
	NB noise – Ipsi, Reflex	1000, 2000, 3000, 4000 Hz
	Stimulus duration:	1.0 sec
	Reflex Acceptance	Adjustable between 2% and 6%, or 0.05 – 0.15 ml change of ear canal volume.
	Intensity max	90, 100, 120, dBHL.
Outputs:	Contra Earphone:	TDH39 earphone, DD45 earphone, CIR33 insert and/or EARTone 3A insert for Reflex measurements.
	Ipsi Earphone:	Probe earphone incorporated in the probe system for Reflex measurements.
	Probe connection	Connection of the electrical and air system to the probe.
Test types:	Manual Reflex	Manual control of all functions.
	Automated Reflex	Single intensities Reflex growth
	Reflex Decay	Automatic, 10 dB above threshold and manually controlled with stimulus durations of 10 .
	Reflex latency	Automated, first 300 ms from stimulus start.
Audiometer Functions		
Signals:	Frequencies Hz:	Intensities dBHL:
	125	-10 to 70
	250	-10 to 90
	500	-10 to 100
	1000	-10 to 100
	2000	-10 to 100
	3000	-10 to 100
	4000	-10 to 100
	6000	-10 to 100
	8000	-10 to 90
Test types	Auto Threshold Determination (Modified Hughson Westlake).	
General		
PC control:	USB:	Input/output for computer communication. AT235 can be fully operated from a PC. The measurements can then be followed on the PC screen. Data can be sent to and saved on the PC and stored in OtoAccess™.
Thermal printer (Optional):	Type: MPT-III	Thermal MPT-III printer with recording paper in rolls. Print on command via USB
Power supply 	UE60	Use only specified power supply unit UE60 type Input: 100-240VAC 50-60Hz, 1.5 A Output: 24.0 VDC
Dimensions	DxWxH: 29x38x7,5 cm	
AT235 Weight	2.5 kg	

Table 1: Frequencies and Intensity Ranges

AT235 Maximums IMP										
	TDH39		CIR33		EARTone 3A / IP30		IPSI		DD45	
Center	Reading		Reading		Reading		Reading		Reading	
Freq.	Tone	NB	Tone	NB	Tone	NB	Tone	NB	Tone	NB
[Hz]	[dB HL]	[dB HL]	[dB HL]	[dB HL]	[dB HL]	[dB HL]	[dB HL]	[dB HL]	[dB HL]	[dB HL]
125	80	65	90	70	85	85	70	60	75	60
250	100	85	100	85	105	100	85	75	100	85
500	115	100	110	100	110	105	100	85	115	100
750	120	105	110	105	120	110	100	85	120	105
1000	120	105	115	105	120	110	105	90	120	105
1500	120	105	115	105	120	110	110	90	120	105
2000	120	105	115	105	120	110	105	90	115	105
3000	120	105	115	105	120	110	95	90	120	105
4000	120	105	110	100	120	105	100	85	115	105
6000	120	100	95	95	105	100	85	80	110	90
8000	105	95	75	80	90	85	80	75	105	95
10000										
WB	-	115	-	115	-	115	-	95	-	120
LP	-	120	-	115	-	120	-	100	-	120
HP	-	115	-	115	-	120	-	195	-	120

Specification of input/output connections			
Inputs	Connector type	Electrical properties	
Patient response	Jack 6.3mm	Handheld switch: Pin 1: GND Pin 2: Signal	
Outputs:			
Phones, Left	Jack, 6.3mm	Voltage: Min. load impedance: Pin 1: CH1 GND Pin 2: CH1 OUT	
Phones, Right	Jack 6.3mm	Pin 1: CH1 GND Pin 2: CH1 OUT	
Phones, Contralateral	Jack 6.3mm	Voltage: Min. load impedance: Pin 1: CH1 GND Pin 2: CH1 OUT	
Probe system	15-pin D-sub highdensity with air connection	Pin 1	DSP_I2C_INTERRUPT
		Pin 2	GND
		Pin 3	IPSI_OUT
		Pin 4	GND_CONTRA
		Pin 5	GND_Probe-mic
		Pin 6	DSP_I2C_SCLK
		Pin 7	GND
		Pin 8	GND_IPSI
		Pin 9	PROBETONE_OUT
		Pin 10	MIC-IN
		Pin 11	DSP_I2C_DATA
		Pin 12	+5 Vprobe
		Pin 13	CONTRA_OUT
		Pin 14	GND_PROBETONE
		Pin 15	MIC-+IN
Data I/O:			
USB	USB type "B"	USB port for communication	

Calibration Properties

Calibrated Transducers:	Contralateral Earphone:	Telephonics TDH39/DD45 with a static force of 4.5N 0.5N and/or EARTone 3A and/or CIR33 insert phone
	Probe system:	Ipsilateral Earphone: is integrated in the probe system
		Probe frequency transmitter and receiver and pressure transducer is integrated in the probe system
Accuracy:	General	Generally the instrument is made and calibrated to be within and better than the tolerances required in the specified standards:
	Reflex Frequencies:	1%
	Contralateral Reflex and Audiometer Tone Levels:	3 dB for 250 to 4000Hz and 5 dB for 6000 to 8000Hz
	Ipsilateral Reflex Tone Levels:	5 dB for 500 to 2000Hz and +5/-10 dB for 3000 to 4000Hz
	Pressure measurement : Compliance measurement:	5% or 10 daPa, whichever is greater 5% or 0.1 ml, whichever is greater
Stimulus Presentation Control:	Reflexes:	ON-OFF ratio = ≥ 70 dB Rise time = 20 ms Fall time = 20 ms A weighted SPL in Off = 31 dB

Impedance Calibration Properties

Probe tone	Frequencies:	226 Hz 1%, 678 Hz 1%, 800 Hz 1%, 1000 Hz 1%
	Level:	85 dB SPL 1.5 dB measured in an IEC 60318-5 acoustic coupler. The level is constant for all volumes in the measurement range.
	Distortion:	Max 1% THD
Compliance	Range:	0.1 to 8.0 ml
	Temperature dependence:	-0.003 ml/C
	Pressure dependence:	-0.00020 ml/daPa
	Reflex sensitivity: Reflex artifact level:	0.001 ml is the lowest detectable volume change ≥ 95 dB SPL (measured in the 711 coupler, 0.2 ml, 0.5 ml, 2.0 ml & 5.0 ml hardwalled cavities).
	Temporal reflex characteristics:	Initial latency = 35 ms (5 ms) Rise time = 42 ms (5 ms) Terminal latency = 23 ms (5 ms) Fall time = 44 ms (5 ms) Overshoot = max. 1% Undershoot = max 1%
Pressure	Range:	Values between -600 to +400 daPa can be selected in the setup.
	Safety limits:	-750 daPa and +550 daPa, 50 daPa

Reflex Calibration Standards and Spectral Properties:

General	Specifications for stimulus and audiometer signals are made to follow IEC 60645-5	
Contralateral Earphone	Pure tone:	ISO 389-1 for TDH39 and ISO 389-2 for CIR 55.
	Wide Band noise (WB): Spectral properties:	Interacoustics Standard As "Broad band noise" specified in IEC 60645-5, but with 500 Hz as lower cut-off frequency.
	Low Pass noise (LP): Spectral properties:	Interacoustics Standard Uniform from 500 Hz to 1600 Hz, 5 dB re. 1000 Hz level
	High Pass noise (HP): Spectral properties:	Interacoustics Standard Uniform from 1600 Hz to 10KHz, 5 dB re. 1000 Hz level
Ipsilateral Earphone	Pure tone:	Interacoustics Standard.

	Wide Band noise (WB): Spectral properties:	Interacoustics Standard As "Broad band noise" specified in IEC 60645-5, but with 500 Hz as lower cut-off frequency.
	Low Pass noise (LP): Spectral properties:	Interacoustics Standard Uniform from 500 Hz to 1600 Hz, 10 dB re. 1000 Hz level
	High Pass noise (HP): Spectral properties:	Interacoustics Standard Uniform from 1600 Hz to 4000 Hz, 10 dB re. 1000 Hz level
	General about levels:	The actual sound pressure level at the eardrum will depend on the volume of the ear. See Table 2 for details.
The risk of artifacts at higher stimulus levels in reflex measurements are minor and will not activate the reflex detection system		

Table 3: Reference Values for Stimulus Calibration

Freq.	Reference Equivalent Threshold Sound Level (RETSPL) [dB re. 20 µPa]						Variation of Ipsi stimulus levels for different volumes of the ear canal Relative to the calibration performed on an IEC 126 coupler [dB]		Sound attenuation values for TDH39/DD45 earphones using MX41/AR or PN51 cushion [dB]
	ISO 389-1 (Interacoustics Standard)	ISO 389-2 (Interacoustics Standard)	ISO 382-2 (Interacoustics Standard)	Interacoustics Standard	Interacoustics Standard	ISO 389-4 (ISO 8798)	0.5 ml	1 ml	
[Hz]	TDH39	EARtone 3A / IP30	CIR55	DD45	Probe	NB Stimulus Correction Values			
125	45	26	26	47.5	41	4			3
250	25.5	14	14	27	24.5	4			5
500	11.5	5.5	5.5	13	9.5	4	9.7	5.3	7
1000	7	0	0	6	6.5	6	9.7	5.3	15
1500	6.5	2	2	8	5	6			21 (1600 Hz)
2000	9	3	3	8	12	6	11.7	3.9	26
3000	10	3.5	3.5	8	11	6	-0.8	-0.5	31 (3150 Hz)
4000	9.5	5.5	5.5	9	3.5	5	-1.6	-0.8	32
6000	15.5	2	2	20.5	3	5			26 (6300 Hz)
8000	13	0	0	12	-5	5			24
RETSPL WB	-8	-5	-5	-8	-5		7.5	3.2	
LP	-6	-7	-7	-6	-7		8.0	3.6	
HP	-10	-8	-8	-10	-8		3.9	1.4	

*All figures in bold are Interacoustics Standard values.